



Certificate of Conformity

1-2/F., South Block, Building A2, No.3, Keyan Road, Science City, Guangzhou High-Tech Industrial Development Zone, Guangdong, China
Te I: 86-20-32209330
www.i-testlab.com

No. ITL-211014006

The following products have been tested by us with the listed standards and found in compliance with the European directive 2014/35/EU. It is possible to use CE marking to demonstrate the compliance with the LVD directive.

Applicant : **Guangdong Biolight Meditech Co., Ltd.**
Address : No.2 Innovation First Road, Technical Innovation Coast, Hi-tech Zone, Zhuhai, P.R. China
Product name : **Veterinary monitor**
Model No. : **M850Vet**
Technical data : For power adapter: input: 100-240V~, 50/60Hz, Max. 0. 5A,
output: 5.0V---2.0A
For Veterinary monitor: 5.0V---2.0A
Test Report No. : 211014006

Test Standards	
EN 60335-1:2012 + AC:2014 + A11:2014 + A13:2017 + A1:2019 + A14:2019 + A2:2019	Household and similar electrical appliances - Safety -- Part 1: General requirements
EN 62233:2008 + AC:2008	Measurement methods for electromagnetic fields of household appliances and similar apparatus with regard to human exposure



The test report was carried out from the submitted type samples of a product in compliance with the specification of the respective standards. The certificate holder has the right to fix the CE-mark for LVD directive on the product complying with the inspection samples